

Certificate of Analysis

Bichichoninic Acid (BCA) Protein Assay Reagent A

Cat No : P8101

Lot No : 07271819

TEST	Specification	Result
Appearance	Clear, colorless to light yellow	PASS
pH	11.25 + 0.2	11.35
A(562) Blank OD @ 562nm	< 0.100	0.059
Slope	Confirm	0.0010
Correlation Coefficient	Confirm	1.00

This document has been produced electronically and is valid without signature

Date of Print : 8. 2. 2018